



Reference values and limits of satisfactory results

MPS-SAA-4/2016

Parameters at the concentration level of drinking and surface water				
Parameter	Unit	Reference value	LL	UL
Ag	µg/l	28,0	22,4	33,6
Al	µg/l	165	132	198
As	µg/l	11,0	8,80	13,2
B	µg/l	252	202	302
Cd	µg/l	0,93	0,74	1,12
Co	µg/l	38,8	31,0	46,6
Cr	µg/l	11,7	9,36	14,0
Cu	µg/l	14,0	11,2	16,8
Fe	µg/l	157	126	188
Mn	µg/l	165	132	198
Ni	µg/l	9,20	7,36	11,0
Pb	µg/l	7,86	6,29	9,43
Sb	µg/l	7,95	6,36	9,54
Zn	µg/l	76,2	61,0	91,4
LL = lower limit of satisfactory results		UL = upper limit of satisfactory results		

Responsible person: RNDr. Jana Tkáčová, PhD.

Bratislava: 13th June 2016