



Reference values and limits of satisfactory results

MPS-SAA-10/2017

Parameters at the concentration level of waste water				
Parameter	Unit	Reference value	LL	UL
Ag	mg/l	0,306	0,245	0,367
Al	mg/l	2,14	1,71	2,57
Ba	mg/l	3,35	2,68	4,02
As	mg/l	0,701	0,561	0,841
Cd	mg/l	0,574	0,459	0,689
Co	mg/l	1,44	1,15	1,73
Cr	mg/l	1,51	1,21	1,81
Cu	mg/l	0,858	0,686	1,03
Fe	mg/l	2,17	1,74	2,60
Mn	mg/l	3,08	2,46	3,70
Ni	mg/l	0,734	0,587	0,881
Pb	mg/l	0,854	0,683	1,02
Se	mg/l	0,576	0,461	0,691
V	mg/l	1,43	1,14	1,72
Zn	mg/l	2,25	1,80	2,70
LL = lower limit of satisfactory results			UL = upper limit of satisfactory results	

Responsible person: Ing. Monika Vicenová

Bratislava, 6th December 2017