



Referenčné hodnoty a interval vyhovujúcich výsledkov

**MPS-SOA-10/2018**

Ukazovatele na koncentračnej úrovni pitnej a povrchovej vody				
Ukazovateľ	Jednotka	Referenčná hodnota	DM	HM
PCB 28	ng/l	50	35,0	65,0
PCB 52	ng/l	40	28,0	52,0
PCB 101	ng/l	20	14,0	26,0
PCB 118	ng/l	50	35,0	65,0
PCB138	ng/l	20	14,0	26,0
PCB 153	ng/l	35	24,5	45,5
PCB 180	ng/l	25	17,5	32,5
antracén	ng/l	40	28,0	52,0
benzo(b)fluorantén	ng/l	40	28,0	52,0
benzo(k)fluorantén	ng/l	72	50,4	93,6
benzo(a)pyrén	ng/l	56	39,2	72,8
benzo(ghi)perylén	ng/l	56	39,2	72,8
fluorantén	ng/l	56	39,2	72,8
fenantrén	ng/l	40	28,0	52,0
Indeno(1,2,3-cd)pyrén	ng/l	40	28,0	52,0
hexachlórbenzén	ng/l	20	14,0	26,0
lindán	ng/l	35	24,5	45,5
heptachlór	ng/l	30	21,0	39,0
p,p-DDT	ng/l	40	28,0	52,0
metoxychlór	ng/l	75	52,5	97,5

DM - dolná medza vyhovujúcich výsledkov      HM - horná medza vyhovujúcich výsledkov

Vypracoval: Ing.Katarína Šimovičová, PhD.

Bratislava: 28.11. 2018