



Referenčné hodnoty a interval vyhovujúcich výsledkov

**MPS-SOA-10/2021**

Ukazovatele na koncentračnej úrovni pitnej/povrchovej/odpadovej vody				
Ukazovateľ	Jednotka	Referenčná hodnota	DM	HM
PCB 28	ng/l	25,0	17,5	32,5
PCB 52	ng/l	30,0	21,0	39,0
PCB 101	ng/l	45,0	31,5	58,5
PCB 118	ng/l	20,0	14,0	26,0
PCB 138	ng/l	30,0	21,0	39,0
PCB 153	ng/l	50,0	35,0	65,0
PCB 180	ng/l	25,0	17,5	32,5
antracén	ng/l	30,0	21,0	39,0
benzo(b)fluorantén	ng/l	30,0	21,0	39,0
benzo(k)fluorantén	ng/l	20,0	14,0	26,0
benzo(a)pyrén	ng/l	40,0	28,0	52,0
benzo(ghi)perylén	ng/l	20,0	14,0	26,0
fluorantén	ng/l	40,0	28,0	52,0
fenantrén	ng/l	30,0	21,0	39,0
indeno(1,2,3-cd)pyrén	ng/l	40,0	28,0	52,0
Hexachlórbenzén	ng/l	35,0	24,5	45,5
Lindán	ng/l	65,0	45,5	84,5
Heptachlór	ng/l	50,0	35,0	65,0
p,p- DDT	ng/l	28,0	19,6	36,4
Metoxychlór	ng/l	45,0	31,5	58,5
DM - dolná medza vyhovujúcich výsledkov			HM - horná medza vyhovujúcich výsledkov	

Vypracoval: Ing.Katarína Šimovičová, PhD.

Bratislava, 24.11.2021